

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 21 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130116</u> ✓	
A	Name <u>Kenneth Parker</u>	ID# <u>5578</u>	Date <u>9-17-09</u>
	Agency <u>McGrath Public Safety</u>	Phone # <u>524-3075</u>	
	Instrument Location <u>AST McGrath</u> ✓		
B	Alco S/N <u>77111</u> ✓	Target Value <u>.078</u> ✓	High Pressure <u>700</u>
	Alco Test Values	<u>.068</u> ✓ 1 <sup>st</sup> Alco	<u>.070</u> ✓ 2 <sup>nd</sup> Alco
	Signature <u>[Signature]</u> ✓		

(OVER)

BMB  
10/13/09

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 20<sup>th</sup> day of October 2009.

(Notary Seal Stamp)

Carolyn M. Noland  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



**VERIFICATION OF CALIBRATION REPORT***of DataMaster cdm breath Test Instrument**State of Alaska**Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*DataMaster cdm S/N 130116 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

**NONDRINKING SUBJECT TEST**ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130116

SEPTEMBER 17, 2009

OPERATOR'S NAME:

PARKER/KENNETH/W

OPERATOR'S NUMBER: 5578

SUBJECT'S LAST NAME:

PARKER

SUBJECT'S FIRST NAME/MI :

KENNETH/W

O.L. #: 130116

DEPT/AGENCY: MC61

CASE/REPORT: T

TEST TYPE: T

ALCO TARGET VALUE: 0.78

ALCO S/N: 77111

**BREATH ANALYSIS**

.078 ADJUSTED FOR 29.18 in

ALCO TARGET .076 11:10

BLANK TEST .000 11:11

INTERNAL STANDARD VERIFIED 11:11

ALCO TV 29.18 in .068 11:12

BLANK TEST .000 11:13

SUBJECT SAMPLE .000 11:13

BLANK TEST .000 11:14

ALCO TV 29.20 in .070 11:15

BLANK TEST .000 11:16

**DIAGNOSTIC CHECK**ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130116

SEPTEMBER 17, 2009 ✓

TIME 11:06

**DIAGNOSTIC CHECK**

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 42c

BAROMETER: 29.18 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

**PRINTER TEST**!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNOPQRSTUVWXYZ\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~■